

ARC BRASOV
ing. Dragos Rotaru
Fantanii str. 17
500482 Brasov
Romania

Our reference number:

CMI-2376/2020/0318

Contact person:

Mgr. Blanka Šmardová, Ph.D.

Brno

Date:

30.03.2020

Final evaluations of the bilateral interlaboratory comparisons

Your laboratory has participated in the following bilateral interlaboratory comparisons:

BILC 179-19

THERMAL IMAGING CAMERA

BILC 180-19

INFRARED THERMOMETER

After evaluation of the measurement results stated in your Calibration Certificates according to EN ISO/IEC 17043 document we can confirm that the differences of your measurement results from the reference values are in conformance with your declared uncertainties and your laboratory herewith fulfilled the conditions stated for this BILC programs and

has proved its technical competence.

This conclusion can be appealed in writing within 15 days of receipt of this Decision.

Enclosed we are sending the final Reports and Certificates on successful participations in these BILC programs.

Thank you for your cooperation during the BILC processes.

Yours sincerely



Ing. František Staněk, PhD.
Director for Legal Metrology

Český metrologický institut

Okružní 31
638 00 Brno

-9-

Annexes:	BILC final report	2 x
	Certificate on successful participation in the BILC program	2 x
Copy:	Legal Metrology / ILC Department	



CZECH METROLOGY INSTITUTE

Český metrologický institut, Referát MPZ



CERTIFICATE

ON PARTICIPATION IN THE BILATERAL INTERLABORATORY COMPARISON

No. 0318-OV-A179-19

Subject of the BILC:

THERMAL IMAGING CAMERA

BILC identification: **32-179-19**

Participant:

ARC BRASOV SRL

Str. Fantanii no. 17, 500482 Brasov, Romania

Final statement

On the basis of the Report No. 0318-ZV-A179-19 on evaluation of the measurement results the above-mentioned participant

c o m p l i e s w i t h

the requirements of this interlaboratory comparison based on EN ISO/IEC 17043.

Date of issue: **30.03.2020**

Mgr. Blanka Šmardová, Ph.D.
Employee of the ILC Department



Ing. František Staněk, PhD.
Director for Legal Metrology



CZECH METROLOGY INSTITUTE

Český metrologický institut, Referát MPZ



CERTIFICATE

ON PARTICIPATION IN THE BILATERAL INTERLABORATORY COMPARISON

No. 0318-OV-A180-19

Subject of the BILC:

INFRARED THERMOMETER

BILC identification: **32-180-19**

Participant:

ARC BRASOV SRL

Str. Fantanii no. 17, 500482 Brasov, Romania

Final statement

On the basis of the Report No. 0318-ZV-A180-19 on evaluation of the measurement results the above-mentioned participant

c o m p l i e s w i t h

the requirements of this interlaboratory comparison based on EN ISO/IEC 17043.

Date of issue: **30.03.2020**

Mgr. Blanka Šmardová, Ph.D.
Employee of the ILC Department



Ing. František Staněk, PhD.
Director for Legal Metrology