

# Supplement

Title: 971 Users Supplement Issue: 4  
Part Number: 2441047 Issue Date: 9/11  
Print Date: Sept. 2005 Page Count: 3  
Revision/Date: 1, 5/06

---

---

This supplement contains information necessary to ensure the accuracy of the above Users Manual.

## Change #1, 40313

On page 8, under **Specifications**, add the following:


Instrument for indoor use only.

On page 9, under **Specifications**, in the **Environment** section, add the following:

**Altitude:** Up to 2000 m

## Change #2


On page 3, following the Note, replace the first paragraph with the following:

Press  to turn on the Meter and start taking measurements.

## Change #3


On page 1, under Introduction delete the **Caution**.

Prior to page 1, replace the Figure with the following:



bdm003f.eps

On page 7, replace Figure 1 with the following:



**Figure 1**

bdmoo2f.eps