

# ELEKTROTECHNICKÝ ZKUŠEBNÍ ÚSTAV



ELECTROTECHNICAL TESTING INSTITUTE - CZECH REPUBLIC  
ELEKTROTECHNISCHE PRÜFANSTALT - TSCHECHISCHE REPUBLIK  
INSTITUT ELECTROTECHNIQUE D'ESSAIS - RÉPUBLIQUE TCHÈQUE  
ЭЛЕКТРОТЕХНИЧЕСКИЙ ИСПЫТАТЕЛЬНЫЙ ИНСТИТУТ - ЧЕШСКАЯ РЕСПУБЛИКА

Pod Lisem 129, 171 02 Praha 8 - Troja

## CERTIFICATE

No.: 1130571

**Product:** Active flexible sensor for measurement of alternating current

**Type:** AMOS M

**Rating:** 230 V AC, current ranges 10; 30; 100; 300 or 1000 A, safety class II, overvoltage category IV,  
supply voltage  $\pm 5 - 12$  V DC, output voltage 1,0 V AC at Inom

**Ordering firm:** MEgA - Měřicí Energetické Aparáty, a.s.  
Česká 390, 664 31 Česká, Czech Republic

**Manufacturer:** MEgA - Měřicí Energetické Aparáty, a.s.  
Česká 390, 664 31 Česká, Czech Republic

**Trade mark:**

**The test results are stated in the test-report No.:** 302970-01/01 of: 11.07.2013, 302970-01/02 of: 12.07.2013

**A sample of the product was found to be in conformity with:**  
EN 61010-1:10, EN 61010-2-032:02  
EN 61326-1:13

**The validity of the certificate is limited to:** 31.7.2016

12.7.2013

Prague

Jarmil Mikulík  
Deputy Head of Certification Body



Stamp



302970-01