

## **Barcodeleser mit RS232-Schnittstelle** **Barcode Scanner with RS232 Interface**

3-349-338-15

3/4.12

---

Die Barcodeleser B3261, Z720A und Z502F sind für folgende Barcodes konfiguriert:  
**EAN13 / CODE 39 / CODE 128**

Der Barcodeleser ist sofort am SECUTEST..., MINITEST MASTER/3P MASTER, SECULIFE ST, Serie PROFITEST MASTER bzw. am SECUTEST (P)SI betriebsbereit. Für den Betrieb am PROFITEST (P)SI-BC bzw. PROFITEST 204 muss der Barcodeleser mit den Codes der jeweiligen Anleitung konfiguriert werden. Für diese Prüfgeräte ist nur der Code 128 möglich.

Möchten Sie andere Codes für den SECUTEST... verwenden, so wenden Sie sich an unseren Produktsupport:

Tel. D 0900 1 8602-00

A/CH +49 911 8602-0 • E-Mail: [support@gossenmetrawatt.com](mailto:support@gossenmetrawatt.com)

Barcode scanner B3261, Z720A and Z502F have been configured for the following barcodes: **EAN13 / CODE 39 / CODE 128**

The barcode scanner is immediately ready for operation when connected with a SECUTEST..., MINITEST MASTER/3P MASTER, SECULIFE ST, Series PROFITEST MASTER or SECUTEST (P)SI instrument.

For operation with PROFITEST (P)SI-BC or PROFITEST 204 the barcode scanner must be configured with the codes indicated in the respective operating instructions. For these testers only code 128 is possible.

Please contact our product support if you wish to use other codes for SECUTEST...: Phone: +49 911 8602-0 • E-Mail: [support@gossenmetrawatt.com](mailto:support@gossenmetrawatt.com)

---

GMC-I Messtechnik GmbH  
Südwestpark 15  
90449 Nürnberg • Germany  
Telefon +49 911 8602-111  
Fax +49 911-8602-777  
E-Mail [info@gossenmetrawatt.com](mailto:info@gossenmetrawatt.com)  
[www.gossenmetrawatt.com](http://www.gossenmetrawatt.com)

Erstellt in Deutschland • Änderungen vorbehalten  
Edited in Germany • Subject to change without notice

 **GOSSEN METRAWATT**